

ISO 6145-7:2018

 **NBN**

EN ISO 6145-7:2018

NBN EN ISO 6145-7:2019

 

Gas analysis - Preparation of calibration gas mixtures using dynamic methods - Part 7: Thermal mass-flow controllers (ISO 6145-7:2018)

Valid from 30-01-2019

Replaces NBN EN ISO 6145-7:2011

ICS: 71.040.40

EUROPEAN STANDARD
NORME EUROPÉENNE
EUROPÄISCHE NORM

EN ISO 6145-7

December 2018

ICS 71.040.40

Supersedes EN ISO 6145-7:2010

English Version

**Gas analysis - Preparation of calibration gas mixtures
using dynamic methods - Part 7: Thermal mass-flow
controllers (ISO 6145-7:2018)**

Analyse des gaz - Préparation des mélanges de gaz
pour étalonnage à l'aide de méthodes dynamiques -
Partie 7: Régulateurs thermiques de débit massique
(ISO 6145-7:2018)

Gasanalyse - Herstellung von Kalibriergasgemischen
mit Hilfe von dynamisch-volumetrischen Verfahren -
Teil 7: Thermische Massendurchflussregler (ISO 6145-
7:2018)

This European Standard was approved by CEN on 16 October 2018.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

EN ISO 6145-7:2018 (E)

Contents Page

European foreword..... 3

European foreword

This document (EN ISO 6145-7:2018) has been prepared by Technical Committee ISO/TC 158 "Analysis of gases" in collaboration with CCMC.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by June 2019, and conflicting national standards shall be withdrawn at the latest by June 2019.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 6145-7:2010.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

The text of ISO 6145-7:2018 has been approved by CEN as EN ISO 6145-7:2018 without any modification.

**Gas analysis — Preparation of
calibration gas mixtures using
dynamic methods —**

**Part 7:
Thermal mass-flow controllers**

*Analyse des gaz — Préparation des mélanges de gaz pour étalonnage
à l'aide de méthodes dynamiques —*

Partie 7: Régulateurs thermiques de débit massique



ISO 6145-7:2018(E)



COPYRIGHT PROTECTED DOCUMENT

© ISO 2018

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Fax: +41 22 749 09 47
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

	Page
Foreword	iv
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Symbols	2
5 Principle	2
6 Set-up	2
6.1 General	2
6.2 Thermal mass-flow controller using a constant current supply	3
6.3 Thermal mass-flow controller under constant temperature control	3
7 Preparation of gas mixtures	4
7.1 Description of the experimental procedure	4
7.2 Range of validity	6
7.3 Operating conditions	6
8 Calculations	7
8.1 Volume fraction	7
8.2 Sources of uncertainty	7
8.3 Uncertainty of measurement	8
Annex A (informative) Pre-mixed gases for the preparation of mixtures of high dilution	9
Annex B (informative) Practical hints	10
Annex C (informative) Calculation of uncertainties	12
Bibliography	14

ISO 6145-7:2018(E)

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 158, *Analysis of gases*.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

This third edition cancels and replaces the second edition (ISO 6145-7:2009), which has been technically revised. The main changes compared to the previous edition are as follows:

- correction of some errors in the formulae in [Annexes A](#) and [C](#);
- minor editorial corrections.

A list of all parts in the ISO 6145 series can be found on the ISO website.

Gas analysis — Preparation of calibration gas mixtures using dynamic methods —

Part 7: Thermal mass-flow controllers

1 Scope

ISO 6145 is a series of documents dealing with various dynamic methods used for the preparation of calibration gas mixtures. This document specifies a method for continuous preparation of calibration gas mixtures, from nominally pure gases or gas mixtures by use of thermal mass-flow controllers. The method is applicable to preparation of mixtures of non-reacting species, i.e. those which do not react with any material of construction of the flow path in the thermal mass-flow controller or the ancillary equipment.

If this method is employed for preparation of calibration gas mixtures the optimum performance is as follows: the relative expanded measurement uncertainty U , obtained by multiplying the standard uncertainty by a coverage factor $k = 2$, is not greater than 2 %.

If pre-mixed gases are used instead of pure gases, mole fractions below 10^{-6} can be obtained. The measurement of mass flow is not absolute and the flow controller requires independent calibration.

The merits of the method are that a large quantity of the calibration gas mixture can be prepared on a continuous basis and that multi-component mixtures can be prepared as readily as binary mixtures if the appropriate number of thermal mass-flow controllers is utilized.

NOTE Gas blending systems, based upon thermal mass-flow controllers, and some including the facility of computerization and automatic control, are commercially available.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 6143, *Gas analysis — Comparison methods for determining and checking the composition of calibration gas mixtures*

ISO 6145-1, *Gas analysis — Preparation of calibration gas mixtures using dynamic volumetric methods — Part 1: Methods of calibration*

ISO 7504, *Gas analysis — Vocabulary*

ISO 12963, *Gas analysis — Comparison methods for the determination of the composition of gas mixtures based on one- and two-point calibration*

ISO 19229, *Gas analysis — Purity analysis and the treatment of purity data*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 7504 apply.