

Geregistreeerde Belgische norm

NBN EN 13205-2

1e uitg., juli 2014

Normklasse: T 96

Blootstelling op de werkplek - Prestatiebeoordeling van een monsternemingsapparaat voor het meten van de concentratie van deeltjes in lucht - Deel 2: Prestatiebeproeving in het laboratorium gebaseerd op de bepaling van de bemonsteringsefficiëntie

Exposition sur les lieux de travail - Évaluation des performances des dispositifs de prélèvement pour le mesurage des concentrations de particules en suspension dans l'air - Partie 2: Essai de performances en laboratoire par détermination de l'efficacité de prélèvement

Workplace exposure - Assessment of sampler performance for measurement of airborne particle concentrations - Part 2: Laboratory performance test based on determination of sampling efficiency

Toelating tot publicatie: 25 juli 2014

Vervangt NBN EN 13205 (2002).

Deze Europese norm EN 13205-2:2014 heeft de status van een Belgische norm.

Deze Europese norm bestaat in drie officiële versies (Duits, Engels, Frans).



Bureau voor Normalisatie - Jozef II-straat 40 - 1000 Brussel - België

Tel: +32 2 738 01 12 - Fax: +32 2 733 42 64 - E-mail: info@nbn.be - NBN Online: www.nbn.be
Bank 000-3255621-10 IBAN BE41 0003 2556 2110 BIC BPOTBEB1 BTW BE0880857592

**norme belge
enregistrée**

NBN EN 13205-2

1e éd., juillet 2014

Indice de classement: T 96

Exposition sur les lieux de travail - Évaluation des performances des dispositifs de prélèvement pour le mesurage des concentrations de particules en suspension dans l'air - Partie 2: Essai de performances en laboratoire par détermination de l'efficacité de prélèvement

Blootstelling op de werkplek - Prestatiebeoordeling van een monsternemingsapparaat voor het meten van de concentratie van deeltjes in lucht - Deel 2: Prestatiebeproeving in het laboratorium gebaseerd op de bepaling van de bemonsteringsefficiëntie

Workplace exposure - Assessment of sampler performance for measurement of airborne particle concentrations - Part 2: Laboratory performance test based on determination of sampling efficiency

Autorisation de publication: 25 juillet 2014

Remplace NBN EN 13205 (2002).

La présente norme européenne EN 13205-2:2014 a le statut d'une norme belge.

La présente norme européenne existe en trois versions officielles (allemand, anglais, français).



Bureau de Normalisation - rue Joseph II 40 - 1000 Bruxelles - Belgique

Tél: +32 2 738 01 12 - Fax: +32 2 733 42 64 - E-mail: info@nbn.be - NBN Online: www.nbn.be

Banque 000-3255621-10 IBAN BE41 0003 2556 2110 BIC BPOTBEB1 TVA BE0880857592

English Version

Workplace exposure - Assessment of sampler performance for
measurement of airborne particle concentrations - Part 2:
Laboratory performance test based on determination of sampling
efficiency

Exposition sur les lieux de travail - Évaluation des performances des dispositifs de prélèvement pour le mesurage des concentrations de particules en suspension dans l'air - Partie 2: Essai de performances en laboratoire par détermination de l'efficacité de prélèvement

Exposition am Arbeitsplatz - Beurteilung der Leistungsfähigkeit von Sammlern für die Messung der Konzentration luftgetragener Partikel - Teil 2: Laborprüfung der Leistungsfähigkeit basierend auf der Bestimmung des Probenahmewirkungsgrads

This European Standard was approved by CEN on 7 May 2014.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

Contents		Page
Foreword.....		4
Introduction		6
1 Scope		7
2 Normative references		7
3 Terms and definitions		7
4 Symbols and abbreviations		8
4.1 Symbols		8
4.1.1 Latin		8
4.1.2 Greek.....		10
4.2 Enumerating subscripts.....		10
4.3 Abbreviations		11
5 Principle.....		11
6 Test method.....		11
6.1 General.....		11
6.2 Test conditions		11
6.3 Test variables		12
6.3.1 General.....		12
6.3.2 Particle size		14
6.3.3 Wind speed.....		14
6.3.4 Wind direction.....		14
6.3.5 Aerosol composition		14
6.3.6 Sampled or internally separated mass		14
6.3.7 Aerosol charge.....		14
6.3.8 Specimen variability		15
6.3.9 Excursion from the nominal flow rate		15
6.3.10 Surface treatments		15
7 Experimental requirements		15
8 Calculation of sampler bias and expanded uncertainty		17
8.1 General.....		17
8.2 Determination of the sampling efficiency		18
8.3 Calculation of sampler bias		18
8.3.1 Calculation of the sampled aerosol concentration		18
8.3.2 Calculation of the ideal sampled aerosol concentration		20
8.3.3 Calculation of the sampler bias.....		21
8.4 Calculation of the expanded uncertainty of the sampler		21
8.4.1 General.....		21
8.4.2 Calibration of sampler test system		22
8.4.3 Estimation of sampled concentration		23
8.4.4 Bias relative to the sampling convention.....		23
8.4.5 Individual sampler variability		24
8.4.6 Excursion from the nominal flow rate		24
8.4.7 Combined uncertainty (of measurement)		28
8.4.8 Expanded uncertainty		31
9 Test report		31
9.1 General.....		31

9.2	Testing laboratory details and sponsoring organisation	31
9.3	Description of the candidate sampler	31
9.4	Critical review of sampling process	32
9.5	Laboratory methods used	32
9.6	Details of experimental design	33
9.7	Presentation of experimental results	33
9.8	Data analysis.....	33
9.9	Candidate sampler performance	33
9.10	Report of workplace comparison	33
9.11	Summary and information for the user of the sampler	33
	Bibliography.....	36

Foreword

This document (EN 13205-2:2014) has been prepared by Technical Committee CEN/TC 137 “Assessment of workplace exposure to chemical and biological agents”, the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by December 2014 and conflicting national standards shall be withdrawn at the latest by December 2014.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document together with EN 13205-1, CEN/TR 13205-3, EN 13205-4, EN 13205-5 and EN 13205-6 supersedes EN 13205:2001.

EN 13205, *Workplace exposure — Assessment of sampler performance for measurement of airborne particle concentrations*, consists of the following parts:

- *Part 1: General requirements;*
- *Part 2: Laboratory performance test based on determination of sampling efficiency* (the present document);
- *Part 3: Analysis of sampling efficiency data* [Technical Report];
- *Part 4: Laboratory performance test based on comparison of concentrations;*
- *Part 5: Aerosol sampler performance test and sampler comparison carried out at workplaces;*
- *Part 6: Transport and handling tests.*

Significant technical changes from the previous edition, EN 13205:2001:

- This part of EN 13205 is based on Annex A of the previous edition, EN 13205:2001.
- The scope has been limited to aerosol samplers, and the current version of the standard is not (directly) applicable to other types of aerosol instruments.
- As this is now a standard in its own right, a clause on symbols used has been added. Almost all definitions are now given either in EN 1540, *Workplace exposure — Terminology* or in Part 1 of this standard.
- The method of calculating the uncertainty of a sampler or a measuring procedure has been revised in order to comply with ENV 13005. The concept of “accuracy” is no longer used, instead the concept of “expanded uncertainty” is used.
- The five major sources of uncertainty due to aspects of the sampling performance of an aerosol sampler (calibration of sampler test system, estimation of sampled concentration, bias relative to the sampling convention, individual sampler variability and excursion from nominal flow rate) are described with equations on how to incorporate these uncertainties into the expanded uncertainty of a sampler. CEN/TR 13205-3 gives recommendations how these entities may be calculated from measured sampling efficiency data.
- The list of the particle size distributions (per sampling convention) to be used for the evaluation of sampler performance has been restricted at the lower end to reflect that particles with an aerodynamic

diameter less than 0,5 μm are not sampled due to aerodynamic forces. In the current version, an additional requirement on the size distributions is that at least 84 % of the aerosol mass consists of particles exceeding 0,5 μm .

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Introduction

EN 481 defines sampling conventions for the particle size fractions to be collected from workplace atmospheres in order to assess their impact on human health. Conventions are defined for the inhalable, thoracic and respirable aerosol fractions. These conventions represent target specifications for aerosol samplers, giving the ideal sampling efficiency as a function of particle aerodynamic diameter.

In general, the sampling efficiency of real aerosol samplers will deviate from the target specification, and the aerosol mass collected will therefore differ from that which an ideal sampler would collect. In addition, the behaviour of real samplers is influenced by many factors such as external wind speed. In many cases there is an interaction between the influence factors and fraction of the airborne particle size distribution of the environment in which the sampler is used.

EN 13205 (all parts) enables manufacturers and users of aerosol samplers to adopt a consistent approach to sampler validation, and provide a framework for the assessment of sampler performance with respect to EN 481 and EN 482.

It is the responsibility of the manufacturer of aerosol samplers to inform the user of the sampler performance under the laboratory conditions¹⁾ specified in this part of EN 13205. It is the responsibility of the user to ensure the actual conditions of intended use are within what the manufacturer specifies as acceptable conditions according to the performance test.

1) The inhalable convention is undefined for particle sizes in excess of 100 μm or for wind speeds greater than 4 m/s. The tests required to assess performance are therefore limited to these conditions. If such large particle sizes or wind speeds actually existed at the time of sampling, it is possible that different samplers meeting this document give different results.

1 Scope

This European Standard specifies a laboratory performance test for samplers for the inhalable, thoracic and respirable aerosol fractions, based on determining the sampling efficiency curve of a candidate sampler at a minimum of nine particle sizes. It specifies methods for testing aerosol samplers under prescribed laboratory conditions in order to test whether the performance of a candidate sampler fulfils the requirements of EN 13205-1:2014.

This part of EN 13205 is applicable to all samplers used for the health-related sampling of particles in workplace air.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 1540, *Workplace exposure — Terminology*

EN 13205-1:2014, *Workplace exposure — Assessment of sampler performance for measurement of airborne particle concentrations — Part 1: General requirements*

CEN/TR 13205-3:2014, *Workplace exposure — Assessment of sampler performance for measurement of airborne particle concentrations — Part 3: Analysis of sampling efficiency data*

EN 13205-5:2014, *Workplace exposure — Assessment of sampler performance for measurement of airborne particle concentrations — Part 5: Aerosol sampler performance test and sampler comparison carried out at workplaces*

EN ISO 13137, *Workplace atmospheres — Pumps for personal sampling of chemical and biological agents - Requirements and test methods (ISO 13137)*

3 Terms and definitions

For the purpose of this document, the terms and definitions given in EN 1540, EN 13205-1:2014 and the following apply.

NOTE With regard to EN 1540, in particular, the following terms are used in this document: total airborne particles, respirable fraction, sampling efficiency, static sampler, thoracic fraction, inhalable fraction, measuring procedure, non-random uncertainty, random uncertainty, expanded uncertainty, standard uncertainty, combined standard uncertainty, uncertainty (of measurement), coverage factor and precision.

3.1

relative concentration

concentration expressed as a fraction of the total airborne concentration

3.2

total airborne particle concentration

concentration of aerosol particles present in the air before the particles are affected by the presence of the sampler, or in the case of a personal sampler by the presence of the person wearing the sampler