

Geregistreeerde Belgische norm

NBN EN 13205-5

1e uitg., juli 2014

Normklasse: T 96

Blootstelling op de werkplek - Prestatiebeoordeling van een monsternemingsapparaat voor het meten van de concentratie van deeltjes in lucht - Deel 5: Prestatiebeproeving van een aërosolmonsternemingsapparaat en vergelijking van monsternemingsapparaten op de werkplek

Exposition sur les lieux de travail - Évaluation des performances des dispositifs de prélèvement pour le mesurage des concentrations de particules en suspension dans l'air - Partie 5: Essais de performances des échantillonneurs d'aérosols, réalisés sur les lieux de travail

Workplace exposure - Assessment of sampler performance for measurement of airborne particle concentrations - Part 5: Aerosol sampler performance test and sampler comparison carried out at workplaces

Toelating tot publicatie: 25 juli 2014

Vervangt NBN EN 13205 (2002).

Deze Europese norm EN 13205-5:2014 heeft de status van een Belgische norm.

Deze Europese norm bestaat in drie officiële versies (Duits, Engels, Frans).

***norme belge
enregistrée***

NBN EN 13205-5

1e éd., juillet 2014

Indice de classement: T 96

Exposition sur les lieux de travail - Évaluation des performances des dispositifs de prélèvement pour le mesurage des concentrations de particules en suspension dans l'air - Partie 5: Essais de performances des échantillonneurs d'aérosols, réalisés sur les lieux de travail

Blootstelling op de werkplek - Prestatiebeoordeling van een monsternemingsapparaat voor het meten van de concentratie van deeltjes in lucht - Deel 5: Prestatiebeproeving van een aerosolmonsternemingsapparaat en vergelijking van monsternemingsapparaten op de werkplek

Workplace exposure - Assessment of sampler performance for measurement of airborne particle concentrations - Part 5: Aerosol sampler performance test and sampler comparison carried out at workplaces

Autorisation de publication: 25 juillet 2014

Remplace NBN EN 13205 (2002).

La présente norme européenne EN 13205-5:2014 a le statut d'une norme belge.

La présente norme européenne existe en trois versions officielles (allemand, anglais, français).

English Version

**Workplace exposure - Assessment of sampler performance for
measurement of airborne particle concentrations - Part 5:
Aerosol sampler performance test and sampler comparison
carried out at workplaces**

Exposition sur les lieux de travail - Évaluation des performances des dispositifs de prélèvement pour le mesurage des concentrations de particules en suspension dans l'air - Partie 5: Essais de performances des échantillonneurs d'aérosols, réalisés sur les lieux de travail

Exposition am Arbeitsplatz - Beurteilung der Leistungsfähigkeit von Sammlern für die Messung der Konzentration luftgetragener Partikel - Teil 5: An Arbeitsplätzen durchgeführte Prüfung der Leistungsfähigkeit des Aerosolsammlers und Sammlervergleich

This European Standard was approved by CEN on 7 May 2014.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

Contents	Page
Foreword.....	4
Introduction	6
1 Scope	7
2 Normative references	7
3 Terms and definitions	7
4 Symbols and abbreviations	8
4.1 Symbols	8
4.1.1 Latin	8
4.1.2 Greek.....	10
4.2 Enumerating subscripts.....	10
5 Principle.....	10
6 Test method.....	11
6.1 General.....	11
6.2 Performance test of personal samplers for the inhalable aerosol fraction	11
6.3 Performance test of static samplers.....	11
6.4 Performance test of personal samplers for the respirable or thoracic aerosol fractions.....	11
6.5 Experiments required for the performance test	12
6.5.1 General.....	12
6.5.2 Candidate sampler bias	12
6.5.3 Candidate sampler variability.....	12
6.5.4 Excursion from the nominal flow rate	12
6.5.5 Collected mass or internally separated mass	13
7 Calculation of sampler bias and expanded uncertainty	14
7.1 Requirements	14
7.2 Correction factor.....	14
7.3 Calculation of concentration ratios	14
7.4 Sources of uncertainty (of measurement)	15
7.4.1 General.....	15
7.4.2 Workplace (test) aerosol concentration, as determined using the validated sampler(s)	15
7.4.3 Validated sampler	16
7.4.4 Candidate sampler bias	16
7.4.5 Individual candidate sampler variability	17
7.4.6 Excursion from the nominal flow rate	17
7.4.7 Collected mass or internally separated mass	18
7.5 Combined standard uncertainty.....	21
7.6 Expanded uncertainty	23
8 Periodic validation	23
9 Test report	23
9.1 General.....	23
9.2 Testing laboratory details and sponsoring organisation	24
9.3 Description of the candidate sampler and validated sampler	24
9.4 Critical review of sampling process	24
9.5 Circumstances of field experiment.....	24
9.6 Details of experimental design.....	24
9.7 Data analysis	25

9.8	Performance	25
9.9	Summary and information for the user	25
	Annex A (normative) Procedure for a workplace comparison of a candidate sampler and a validated sampler in order to obtain a correction factor	26
	Bibliography	31

Foreword

This document (EN 13205-5:2014) has been prepared by Technical Committee CEN/TC 137 “Assessment of workplace exposure to chemical and biological agents”, the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by December 2014 and conflicting national standards shall be withdrawn at the latest by December 2014.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document together with EN 13205-1, EN 13205-2, CEN/TR 13205-3, EN 13205-4 and EN 13205-6 supersedes EN 13205:2001.

EN 13205, *Workplace exposure — Assessment of sampler performance for measurement of airborne particle concentrations*, consists of the following parts:

- *Part 1: General requirements;*
- *Part 2: Laboratory performance test based on determination of sampling efficiency;*
- *Part 3: Analysis of sampling efficiency data [Technical Report];*
- *Part 4: Laboratory performance test based on comparison of concentrations;*
- *Part 5: Aerosol sampler performance test and sampler comparison carried out at workplaces (the present document);*
- *Part 6: Transport and handling tests.*

Significant technical changes from the previous edition, EN 13205:2001:

- This part of EN 13205 is partly based on Annex C of the previous edition, EN 13205:2001.
- The scope has been limited to aerosol samplers, and the current version of the standard is not (directly) applicable to other types of aerosol instruments.
- As this is now a standard in its own right, a clause on used symbols has been added. Almost all definitions are now given either in EN 1540, *Workplace exposure — Terminology* or in Part 1 of this standard.
- The method of calculating the uncertainty of a sampler or a measuring procedure has been revised in order to comply with ENV 13005. The concept of “accuracy” is no longer used, instead the concept of “expanded uncertainty” is used.
- The main part of the standard states how to determine the performance of an aerosol sampler at a specific workplace. This is an adaption of the laboratory method given in Part 2.
- The standard gives a method on how to determine the dependence of the sampling efficiency on the collected mass or internally separated mass.

The five major sources of uncertainty due to aspects of the sampling performance of an aerosol sampler (calibration of sampler test system, estimation of sampled concentration, bias relative to the sampling

convention, individual sampler variability and excursion from nominal flow rate) are described with formulae on how to incorporate these uncertainties into the expanded uncertainty of a sampler.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Introduction

EN 481 defines sampling conventions for the particle size fractions to be collected from workplace atmospheres in order to assess their impact on human health. Conventions are defined for the inhalable, thoracic and respirable aerosol fractions. These conventions represent target specifications for aerosol samplers, giving the ideal sampling efficiency as a function of particle aerodynamic diameter.

In general, the sampling efficiency of real aerosol samplers will deviate from the target specification, and the aerosol mass collected will therefore differ from that which an ideal sampler would collect. In addition, the behaviour of real samplers is influenced by many factors such as external wind speed. In many cases there is an interaction between the influence factors and fraction of the airborne particle size distribution of the environment in which the sampler is used.

The workplace performance test for samplers for the inhalable, thoracic or respirable aerosol fractions described in this document is based on a comparison of concentrations sampled from a specific workplace (under otherwise identical conditions) by a candidate sampler and a (previously) validated sampler. Additionally, a method is described for determining a correction factor for recalculation of the concentration determined with one sampler into that of the other at specific workplaces.

This method is intended for the user, rather than the manufacturer, of aerosol samplers.

EN 13205 (all parts) enables manufacturers and users of aerosol samplers to adopt a consistent approach to sampler validation, and provide a framework for the assessment of sampler performance with respect to EN 481 and EN 482.

It is the responsibility of the manufacturer of aerosol samplers to inform the user of the sampler performance under the laboratory conditions¹⁾ specified in EN 13205-1. It is the responsibility of the user to ensure that the actual conditions of intended use are within what the manufacturer specifies as acceptable conditions according to the performance test.

1) The inhalable convention is undefined for particle sizes in excess of 100 µm or for wind speeds greater than 4 m/s. The tests required to assess performance are therefore limited to these conditions. Should such large particle sizes or wind speeds actually exist at the time of sampling, it is possible that different samplers meeting this part of EN 13205 give different results.

1 Scope

This European Standard specifies a method for determining the performance of an aerosol sampler under prescribed workplace conditions in order to test whether the performance of a candidate sampler fulfils the requirements of EN 13205-1.

This part of EN 13205 specifies also a simple method to determine how, for a specific workplace aerosol, the concentration measured by the candidate sampler can be recalculated into that of a validated sampler.

This part of EN 13205 is applicable to all samplers used for the health-related sampling of particles in workplace air. Different test procedures and types of evaluation are included to enable application of this part of EN 13205 to a wide variety of instruments.

The methods specified in this part of EN 13205 are not applicable to tests where the performance of personal samplers is related to static samplers or vice versa.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 1540, *Workplace exposure - Terminology*

EN 13205-1:2014, *Workplace exposure — Assessment of sampler performance for measurement of airborne particle concentrations — Part 1: General requirements*

EN 13205-2:2014, *Workplace exposure — Assessment of sampler performance for measurement of airborne particle concentrations — Part 2: Laboratory performance test based on determination of sampling efficiency*

EN 13205-4:2014, *Workplace exposure — Assessment of sampler performance for measurement of airborne particle concentrations — Part 4: Laboratory performance test based on comparison of concentrations*

3 Terms and definitions

For the purpose of this document, the term and definitions given in EN 1540, EN 13205-1:2014 EN 13205-2:2014 and the following apply.

NOTE With regard to EN 1540, in particular, the following terms are used in this document: total airborne particles, respirable fraction, sampling efficiency, static sampler, thoracic fraction, inhalable fraction, measuring procedure, non-random uncertainty, random uncertainty, expanded uncertainty, standard uncertainty, combined standard uncertainty, uncertainty (of measurement), coverage factor, precision and analysis.

3.1

correction function

mathematical function relating aerosol concentrations measured using a candidate sampler to those measured using a validated sampler, determined by a comparison of the two samplers